



Certificate Number  
QPS19ATEX0002X

**CERTIFICATE OF COMPLIANCE  
ATEX Essential Health & Safety Requirements  
DIRECTIVE 2014/34/EU**

Issued to	AMETEK Process & Analytical Instruments Division
Address	150 Freeport Rd Pittsburgh, PA 15238 United States of America
Project Number	ATX35755-1
Product	AmeVision Display User Interface
Model Number	AMEVision Display User Interface
Markings	Ex II 3 G Ex ec nC IIC T6 Gc Tamb = -20°C to +60°C.
Ratings	Electrical: 115-230 VAC +/-10%, 47-63 Hz, 40 VA max.
Applicable Standards	For ATEX Directive 2014/34/EU Annex II and: EN 60079-0:2018 , EN 60079-7: 2015, EN 60079-15 :2011
Factory/Manufacturing Location	AMETEK Process & Analytical Instruments Division 150 Freeport Rd Pittsburgh, PA 15238 United States of America
Specific Conditions of Use	<ol style="list-style-type: none"><li>1. Transient protection up to 7 V shall be provided during installation for the Ethernet connections,</li><li>2. The product shall be provided with protection against exposure to daylight and light from luminaires,</li><li>3. Unused entries must be blanked using the appropriate Ex ec blanking plugs.</li><li>4. The cable glands and blanking plugs must also maintain an ingress protection rating minimum of IP54.</li><li>5. The 24 Position Terminal blocks is for low voltage signal connections.</li></ol>

QPS Evaluation Services Inc., Unit #8, 81 Kelfield Street, Toronto, Ontario M9W 5A3, Canada, certifies that this equipment has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of the equipment intended for use in potentially explosive atmospheres given in Annex II to the Directive.

Issued By: Dave Adams. P.Eng.

Signature:

Date: May 15, 2019