

Listing #: E113936  
Report #: 84226 c4 136612 r1  
Original Certification Date: March 25, 2015  
Revised Certification Date: March 5, 2026

This Certification is issued to:  
Ametek Process Instruments  
150 Freeport Road  
Pittsburgh, PA 15238  
USA



Stating that the product(s):  
Tail Gas Analyzer,  
Model 888

Product Rating(s):  
• 120/240 VAC, 50/60 Hz, 500 W  
• Class I, Division 2, Groups A to D, T3

Achieved Certification to the following standard(s):  
UL 61010-1/CSA C22.2 No. 61010-1, Third Edition: Safety of Electrical Equipment for Measurement, Control, and Laboratory Use, Rev. May 11, 2012  
UL 121201, Ninth Edition/CSA C22.2 No. 213-17, Third Edition: Nonincendive Electrical Equipment for Use in Class I and II, Division 2 and Class III, Divisions 1 and 2 Hazardous (Classified) Locations, Rev April 1, 2021



Nehemya Cohen  
Certification Officer,  
Eurofins Electrical and Electronic Testing North America, LLC

*All changes proposed in the previously identified product that affects the above information must be submitted to Eurofins for evaluation prior to implementation to assure continued NRTL Certification status. The covered product(s) shall be subject to follow-up inspections to ensure that the Certified product(s) are identical to the product sample evaluated by Eurofins E&E NA and that all responsibilities are being fulfilled as specified in the Applicants' Responsibility section of the Certification Report. The Applicant named above has been authorized by Eurofins E&E NA to represent the product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the Eurofins E&E NA Applicant Contract, Listing Reports, and the applicable agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in this Certification program. This certification has been granted under a System 3 program as defined in ISO/IEC 17067.*



*Eurofins E&E North America, LLC is recognized by OSHA and accredited by the Standards Council of Canada.*

**N**  
**R**  
**T**  
**L**